

STATE OF NEW HAMPSHIRE
BUREAU OF PURCHASE AND PROPERTY
STATE HOUSE ANNEX
25 CAPITOL STREET
CONCORD, NEW HAMPSHIRE 03301-6398

ADDENDUM # 2 TO BID INVITATION # 204-16

DATE OF BID OPENING: 5/3/16

TIME OF BID OPENING: 11:00 AM (EST)

FOR: RAE SYSTEM GAS METERS AND CALIBRATION EQUIPMENT

CHANGE: Change the description on Line 1 of the Response Table to the following:

MultiRae Lite 4 Gas w/ PID

QUESTIONS AND ANSWERS

QUESTION #1

I have a question about line item 4. Is this a case for the monitors or for the calibration gas or both?

ANSWER #1

The case is for the monitors themselves.

The case is to be designed to house 2 MultiRae Lite 4 gas w/PID (Line 1) and two QRAE MultiGas Detector (line 2). It should be a foam filled case which includes a pass-thru charger connector and multi-prong charging adapter. The intent is to have all 4 charges plugged in, inside the case and then plug the case into an outlet to charge the meters.

QUESTION #2

Also, you haven't listed any tubing that would be required for connecting the AutoRAE2 system to the calibration gas/regulators so I'm not sure if you would want to add that or not.

ANSWER #2

It was our understanding the tubing was part of the controller and cradle system. If not, then yes the tubing needs to be included.

PURCHASING AGENT: Robert Lawson
TEL. NO.: 603/271- 3147

NOTE: IN THE EVENT THAT YOUR BID INVITATION HAS BEEN SENT TO THIS OFFICE PRIOR TO RECEIVING THIS ADDENDUM, RETURN ADDENDUM WITHIN THE SPECIFIED TIME WITH ANY CHANGES YOU MAY WISH TO MAKE AND MARK ON THE REMITTANCE ENVELOPE BID INVITATION NUMBER AND OPENING DATE. RETURNED ADDENDA WILL SUPERSEDE PREVIOUSLY SUBMITTED BID.

BIDDER _____ ADDRESS _____

BY _____
(this document must be signed)

_____ TEL. NO. _____
(please type or print name)

Please visit: <http://das.nh.gov/purchasing>
(click on "Bid, Proposals...") for complete bid and addendums.